



Azienda Ospedaliero-Universitaria Careggi - Firenze

Laboratorio di Microbiologia e Virologia
Centro Regionale di Riferimento per la Diagnostica dei Micobatteri
<http://www.aou-careggi.toscana.it/microbiologia/CRRM/>

Result of the sterilisation test carried out during testing of the NocoSystem disinfection appliance

We hereby certify that all the cultures incubated using specimens contaminated with *Mycobacterium tuberculosis*, exposed on work benches and inside level II biosafe environments (in operation) during the testing of the NocoSystem sterilisation apparatus posted a **NEGATIVE** result.

In witness thereof,

(Dr. Enrico Tortoli)

Florence 27 November 2008